

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO. FILING DATE		JNG DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/018,985 12/19/2001		2/19/2001	Olle Inganas	BERGLUNDS P0021	4696	
27667	7590	09/17/2004		EXAM	EXAMINER	
HAYES, SO			DAWSON, GLENN K			
TUCSON, AZ 85701				ART UNIT	PAPER NUMBER	
				3731		

DATE MAILED: 09/17/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

				12 7					
		A	Application No. Applicant(s)						
0.000 - 4 - 4 - 1 - 2 - 2 - 1 - 1 - 1			0/018,985	INGANAS ET A	AL.				
	Office Action Summary	E	xaminer	Art Unit					
			lenn K Dawson	3731					
Period fo	The MAILING DATE of this communi or Reply	cation appear	s on the cover sheet w	vith the correspondence	address				
THE - Exte after - If the - If NC - Failu Any	ORTENED STATUTORY PERIOD FOR MAILING DATE OF THIS COMMUNION IN THIS COMMUNION IN THE PROPERTY OF THIS COMMUNION IN THE PROPERTY OF THE PROPERT	CATION. of 37 CFR 1.136(a) unication. of days, a reply with utory period will a will, by statute, cau	i. In no event, however, may a nin the statutory minimum of th pply and will expire SIX (6) MC se the application to become A	reply be timely filed irty (30) days will be considered ti NTHS from the mailing date of th BANDONED (35 U.S.C. § 133).	is communication.				
Status									
1) 又	Responsive to communication(s) filed	d on <i>01 June</i>	2004.						
• —	2a) This action is FINAL . 2b) This action is non-final.								
,	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.								
Dispositi	ion of Claims								
5)□ 6)⊠ 7)□									
Applicati	on Papers								
10)⊠	The specification is objected to by the The drawing(s) filed on 19 December Applicant may not request that any object Replacement drawing sheet(s) including The oath or declaration is objected to	2001 is/are: tion to the draw the correction	wing(s) be held in abeya is required if the drawin	nce. See 37 CFR 1.85(a) g(s) is objected to. See 37). ' CFR 1.121(d).				
Priority ι	ınder 35 U.S.C. § 119								
12)⊠ a)İ	Acknowledgment is made of a claim f All b) Some * c) None of: 1. Certified copies of the priority of 3. Copies of the certified copies of application from the Internation See the attached detailed Office action	locuments ha locuments ha f the priority nal Bureau (P	ave been received. ave been received in a documents have been CT Rule 17.2(a)).	Application No n received in this Nation	nal Stage				
Attachmen			_						
2) 🔲 Notic 3) 🔯 Inforr	e of References Cited (PTO-892) e of Draftsperson's Patent Drawing Review (PT nation Disclosure Statement(s) (PTO-1449 or F r No(s)/Mail Date <u>03-25, 06-19-2002</u> .		Paper No	Summary (PTO-413) (s)/Mail Date Informal Patent Application (f 	PTO-152)				

Application/Control Number: 10/018,985

Art Unit: 3731

Election/Restrictions

Claims 2-9,13,14,16-21,24-26 and 28-33 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected species, there being no allowable generic or linking claim. Election was made without traverse in the reply filed on 06-01-04.

The applicant should review the claims and the elected species and determine which claims are actually readable on a conjugated polymer layer whose function is implanting; a tool array with identical tools; and the tool being an expandable object.

The examiner is unable to detect any disclosed embodiments of an array of tools on a needle inside a catheter where the tool is an expandable cylindrical object for implanting material into the body.

Drawings

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the tool or tool arrays being on a needle inserted in a catheter must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate

Art Unit: 3731

figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. The replacement sheet(s) should be labeled "Replacement Sheet" in the page header (as per 37 CFR 1.84(c)) so as not to obstruct any portion of the drawing figures. If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1,10-12,15,22,23,27,34 and 35 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

In claim 1, there is no clear antecedent basis for "the tool".

In claim 15, there is no antecedent basis for "the individual tool".

Allowable Subject Matter

Claims 1, 10-12,15,22,23,27,34 and 35 would be allowable if rewritten or amended to overcome the rejection(s) under 35 U.S.C. 112, 2nd paragraph, set forth in this Office action.

Art Unit: 3731

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Glenn K Dawson whose telephone number is 703-308-4304. The examiner can normally be reached on M-Th 7:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhtuan T. Nguyen can be reached on 703-308-2154. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Glenn K Dawson Primary Examiner Art Unit 3731

Gkd 30 August 2004